

Hanna Instruments SRL
Romania, Salaj, Nusfalau, No. 1, Hanna Str.
Tel.: +40/260/607700; Fax: +40/260/607739



Document No: RN-0056

CE DECLARATION OF CONFORMITY

**Hanna Instruments SRL - Nusfalau Science Park
Romania, Salaj, Nusfalau, No. 1, Hanna Str.
Tel.: +40/260/607700; Fax: +40/260/607739**

This certifies that the following designated products

**HI 701, HI 711, HI 713, HI 718
HI 721, HI 736, HI 755, HI 764**

are in conformity with the EMC directive 2004/108/EC and its standards, and Low Voltage Directive 2006/95/EC and its standards:

EN 61326-1: 2006 Electrical Equipment for Measurement, Control and Laboratory use
- EMC requirements Part 1: General requirements
and its standards:

Class A

EN 61000-4-2: 2009, EMC - Electrostatic Discharge Immunity Test

4 kV/4kV contact/air

EN 61000-4-3: 2006, EMC - Radiated RF, Electromagnetic Field Immunity Test

3 V/m (80 MHz to 1 GHz)

EN 61010-1: 2001 (IEC 61010-1: 2001) Safety Requirements for Electrical Equipment
for Measurement, Control and Laboratory Use Part 1: General Requirements.

Date of issue:

September 28, 2010

Name and Function:

Marius Todorutiu elec.eng.

A handwritten signature in blue ink, appearing to read 'Marius', is written over a horizontal line.